K032565 1 of 1

OCT 2 2 2003

510(k) SUMMARY

Submitter:	Cynsoure, Inc. 10 Elizabeth Drive Chelmsford, MA 01824
Contact:	George Cho Senior Vice President of Medical Technology
Date Summary Prepared:	August 19, 2003
Device Trade Name:	PhotoGenica VL
Common Name:	Medical Laser System
Classification Name:	Instrument, surgical, powered, laser 79-GEX 21 CFR 878.4810
Equivalent Device:	Nlite Laser by ICN, and Cynosure PhotoGenica VL
Device Description:	The PhotoGenica VL is a pulse-dye laser, having the organic dye as the lasing medium. It is a pulsed laser with a wavelength of 580 to 590nm. Laser activation is both by finger switch and footswitch. Overall weight of the laser is 285lbs, and the size is 44"x19"x24" (HxWxD). Electrical requirement is 110 VAC or 220 VAC, 20A, 50-60 Hz, single phase.
Intended Use:	The PhotoGenica VL is indicated for treatment of periocular wrinkles, vascular lesions and inflammatory acne.
Comparison:	The PhotoGenica VL laser has an equivalent indication for uses, the same principle of operation, the same wavelength and pulse energy range as the predicate devices.
Nonclinical Performance Data:	none
Clinical Performance Data:	none
Conclusion:	The PhotoGenica VL Laser is another safe and effective device for dermatologic applications.
Additional Information:	none

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

OCT 2 2 2003

Mr. George Cho
Senior Vice President of Medical Technology
Cynosure, Inc.
10 Elizabeth Drive
Chelmsford, Massachusetts 01824

Re: K032565
Trade/Device Name: PhotoGenica VL
Regulation Number: 21 CFR 878.4810
Regulation Name: Laser surgical instrument for use in general and plastic surgery and in dermatology
Regulatory Class: II
Product Code: GEX
Dated: August 19, 2003
Received: August 21, 2003

Dear Mr. Cho:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4659. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,

for Mark N Melkerson

Celia M. Witten, Ph.D., M.D.
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

501(k) Number (if known): K032565

Device Name: Cynosure PhotoGenica VL

Indications For Use:

The Cynosure PhotoGenica VL laser is indicated for use in Dermatological and Plastic Surgery applications and the treatment of periocular wrinkles, vascular lesions and inflammatory acne vulgaris.

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

for Mark N Melkerson

(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K032565

Prescription Use ✓
(Per 21 CFR 801.109)

OR

Over-the-Counter Use _____

(Optional Format 1-2-96)